

INNOVATION

Open Access



An empirical model for educational simulation of cervical dilation in first-stage labor

Silvano R. Gefferie^{1†}, Anouk W. J. Scholten^{1†}, Kim A. E. Wijlens^{1*†} , M. Luísa Ferreira Bastos², M. Beatrijs van der Hout-van der Jagt³, Hans Zwart⁴ and Willem J. van Meurs⁵

Abstract

Background: Several models for educational simulation of labor and delivery were published in the literature and incorporated into a commercially available training simulator (CAE Healthcare Lucina). However, the engine of this simulator does not include a model for the clinically relevant indicators: uterine contraction amplitude and frequency, and cervical dilation. In this paper, such a model is presented for the primigravida in normal labor.

Methods: The conceptual and mathematical models represent oxytocin release by the hypothalamus, oxytocin pharmacokinetics, and oxytocin effect on uterine contractions, cervical dilation, and (positive) feedback from cervical dilation to oxytocin release by the hypothalamus.

Results: Simulation results for cervical dilation are presented, together with target data for a normal primigravida. Corresponding oxytocin concentrations and amplitude and frequency of uterine contractions are also presented.

Conclusion: An original empirical model for educational simulation of oxytocin concentration, uterine contractions, and cervical dilation in first-stage labor is presented. Simulation results for cervical dilation match target data for a normal patient. The model forms a basis for taking into account more independent variables and patient profiles and can thereby considerably expand the range of training scenarios that can be simulated.

Keywords: Model, First stage labor, Oxytocin, Uterine contractions, Cervical dilation, Educational simulation

Introduction

Uterine contractions, fetal descent, and cervical dilation are used to assess progression of labor. These related phenomena vary greatly among parturients and may have an incidence on the condition of the fetus. Simulators allow for practice of normal and critical situations at will and at no risk to real patients. Several models for educational simulation of labor and delivery were published in the literature [1–3] and incorporated into a commercially available training simulator [4]. The cited models make amplitude, frequency, duration, and resting tone of the uterine pressure waveform evolve spontaneously and under the influence of

oxytocin and tocolytics. However, the relationships between these variables and cervical dilation are scripted.

A positive feedback loop involving oxytocin plays a central role in the progression of first-stage labor. The hypothalamus releases oxytocin into the bloodstream, which is transported to the uterus and stimulates uterine contractions. The contractions result in descent of the fetus. The descending fetal head causes cervical dilation. Cervical stretch receptors send a signal to the hypothalamus via afferent nerves, stimulating additional release of oxytocin, thereby closing the loop. This loop results in progressively intensifying contractions and dilation over time [5].

A database for a number of contraction and dilation profiles could be established and used for scripted training simulation purposes. However, realistic simulation of the interaction between oxytocin concentrations, uterine contractions, fetal descent, and dilation, for a number of patient profiles, and possibly under external influences

* Correspondence: k.a.e.wijlens@student.utwente.nl

[†]Silvano R. Gefferie, Anouk W. J. Scholten and Kim A. E. Wijlens contributed equally to this work.

¹Technical Medicine, University of Twente, Drienerloolaan 5, 7522 NB Enschede, The Netherlands

Full list of author information is available at the end of the article



such as administration of exogenous oxytocin and tocolytics, requires a model-based approach. Such a model would not only increase insight in the interacting processes, but also greatly expand the range of scenarios that can be simulated beyond the relatively simple phenomena that can be characterized in a database.

In this innovation paper, we lay the foundation for such a model by proposing an empirical model that simulates oxytocin concentrations and contraction and dilation profiles. The model should be able to run in real, or accelerated, time and should be easy to manipulate to reflect different patients.

Methods

Basic assumptions are made, similar to the ones in the article of Bastos et al. [3] and based on understanding of physiological relations, in order to keep the model as simple as possible, while approaching the target data of Zhang et al. [6]. The subsystems and variables connecting them, as described in the introduction, are reflected in the block diagram of Fig. 1.

The model equations referring to Fig. 1 from left to right are introduced below. The rate of oxytocin release by the hypothalamus $r(t)$ in mU/min has a fixed component, reflected in the parameter P_1 and depends linearly via a gain P_2 on the sensed dilation $d(t)$ in centimeters:

$$r(t) = P_1 + P_2d(t) \tag{1}$$

Units of parameters can easily be derived from the units of the variables. The change in mass of oxytocin depends on the elimination rate constant P_3 , the current mass $m(t)$ in mU, and the release rate, represented by a first-order pharmacokinetic model:

$$\dot{m}(t) = -P_3m(t) + r(t) \tag{2}$$

Dividing the mass by the volume of distribution P_4 results in the concentration of oxytocin in mU/mL:

$$c(t) = \frac{m(t)}{P_4} \tag{3}$$

The concentration dependency of contraction frequency $f(t)$ in 1/min and contraction amplitude $a(t)$ in mm Hg is

represented by a sigmoidal curve, Eqs. (4) and (5). For example in Eq. (4), for a concentration equal to zero, the frequency is zero; for a concentration equal to P_6 , it is 50% of the maximum frequency P_5 ; and for higher concentrations, it tends toward P_5 . The slope of both pharmacodynamic relationships is governed by P_7 :

$$f(t) = P_5 \frac{c(t)^{P_7}}{P_6^{P_7} + c(t)^{P_7}} \tag{4}$$

The amplitude has a non-zero baseline P_8 and a concentration-dependent part reflected in P_9 (the difference between the maximum and baseline amplitude values):

$$a(t) = P_8 + P_9 \frac{c(t)^{P_7}}{P_6^{P_7} + c(t)^{P_7}} \tag{5}$$

The change in dilation depends on the pressure exerted by the fetus on the cervix P_{10} , which is assumed fixed in this version of the model, and on the product of frequency and amplitude. This second influence is assumed linearly proportional with gain P_{11} . Thus, the resulting uterine model is

$$\dot{d}(t) = P_{10} + P_{11}f(t)a(t) \tag{6}$$

Referenced and empirically adjusted parameters are listed in Table 1.

Equations (1–6) were implemented in MATLAB (MATLAB R2017a, The MathWorks Inc.) using the Euler forward method with an integration step size of 1 min to integrate the two differential equations. Model simulation results are presented and compared to target data obtained from Zhang et al. [6] for cervical dilation.

Results

Figure 2 shows simulation results from the initial conditions $d(0) = 2.00$ cm, and $m(0) = 275$ mU, using the parameters listed in Table 1. The stars in Fig. 1 correspond to dilation data for patient A in Zhang et al. [6]. This patient was admitted at 2 cm and labor progressed to 10 cm with labor considered non-protracted.




Fig. 1 Conceptual model for educational simulation of cervical dilation during first-stage labor. Oxt.: oxytocin

Table 1 Model parameters. For consistency, all values are given in three significant digits

Symbol	Description	Value	Unit	Reference
P_1	Baseline oxytocin release rate in hypothalamus	0.740	mU/min	[7]
P_2	Gain from cervical dilation to oxytocin release rate in hypothalamus	50.0	mU/(min cm)	
P_3	Pharmacokinetic parameter: Elimination rate constant	0.0693	1/min	[3]
P_4	Pharmacokinetic parameter: Volume of distribution	18,700	mL	[8, 9]
P_5	Maximal concentration frequency	0.500	1/min	[2]
P_6	Pharmacodynamic parameter: Oxytocin concentration resulting in 50% effect	7.90	mU/mL	[3]
P_7	Pharmacodynamic parameter: Slope of sigmoidal curve	1.11	dimensionless	[2]
P_8	Baseline contraction amplitude	40.0	mm Hg	[10]
P_9	Maximal contraction amplitude	40.0	mm Hg	[2]
P_{10}	Dilation increase due to pressure exerted by the fetus on the cervix	1.00×10^{-3}	cm/min	[11]
P_{11}	Dilation increase due to contraction frequency and amplitude	1.90×10^{-2}	cm/mm Hg	

Discussion

Simulated dilation matches dilation in a normal patient in an approximation that is considered realistic enough for educational simulation. Parameters P_2 and P_{11} were adjusted to achieve this match. Further research could address which parameters are patient- or population-specific and can be used to simulate different labor types. The model parameter estimation procedure could then be formalized so that a clinical instructor can easily use it to meet his or her educational objectives. Ongoing work involves expanding the set of target data and matching simulation results to them, to enhance model validity and demonstrate the possibility to match different patient profiles. Critical analysis of the pharmacokinetic model may be

necessary. It would also be interesting to explore validity of the model for exogenous oxytocin administration.

Conclusion

An original empirical model for educational simulation of oxytocin concentrations, uterine contractions, and cervical dilation in first-stage labor is presented. Simulation results for cervical dilation match target data for a normal patient. The proposed model forms a sound, explicit basis for taking into account more independent variables and patient profiles, and thereby considerably expand the range of training scenarios that can be simulated.




Fig. 2 Simulation results and dilation target data

Acknowledgements

Silvano Gefferie won the Nicolaas Westerhof first prize for a young investigator for his presentation of the work elaborated upon in this paper at the conference Applied Modeling in Acute Care, April 20–21, 2017, Enschede, The Netherlands. Part of the salary of Willem van Meurs in the 2016–2017 academic year was funded by a grant from CAE Healthcare, Montreal, Canada.

Availability of data and materials

The datasets used and/or analyzed during the current study are available from the corresponding author on reasonable request.

Authors' contributions

SG, AS, and KW did the research and wrote the manuscript. All authors read and approved the final manuscript. HZ provided general guidance in the first stages of the research and gave his perspective on the mathematical model in later stages. WvM provided guidance throughout the research project and feedback on academic writing and style. LFB clarified the implementation of her published models for simulation of the effect of oxytocin on uterine contractions. MBvdH-vdJ gave general feedback on the physiological, modeling, and educational simulation aspects of the manuscript.

Ethics approval and consent to participate

Not applicable.

Competing interests

Drs. Ferreira Bastos and Van Meurs received royalties for licensing of the Childbirth simulator mentioned in the introduction by the University of Porto to CAE Healthcare, Montreal, Canada. In the 2016–2017 academic year, Willem van Meurs was a consultant to and received a grant from CAE Healthcare, Montreal, Canada. The remaining authors declare that they have no competing interests.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Author details

¹Technical Medicine, University of Twente, Drienerlolaan 5, 7522 NB Enschede, The Netherlands. ²Animals in Science, Eurogroup for Animals, Hertogstraat 29, 1000 Brussels, Belgium. ³Department of Signal Processing Systems, Faculty of Electrical Engineering, Eindhoven University of Technology, PO Box 513, 5600 MB Eindhoven, The Netherlands. ⁴Department of Applied Mathematics, Faculty of Electrical Engineering, Mathematics, and Computer Science, University of Twente, Drienerlolaan 5, 7522 NB Enschede, The Netherlands. ⁵Cardiovascular and Respiratory Physiology Group, Faculty of Science and Technology, University of Twente, Drienerlolaan 5, 7522 NB Enschede, The Netherlands.

Received: 31 January 2018 Accepted: 21 May 2018

Published online: 15 June 2018

References

- Bastos LF, Lobo MF, van Meurs WL, Ayres-de-Campos D. An intrauterine pressure generator for educational simulation of labour and delivery. *Med Eng Phys.* 2010;32:740–5.
- Bastos LF, van Meurs W, Ayres-de-Campos D. A model for educational simulation of the evolution of uterine contractions during labor. *Comput Methods Prog Biomed.* 2012;107:242–7.
- Lobo MF, Bastos LF, van Meurs WL, Ayres-de-Campos D. A model for educational simulation of the effect of oxytocin on uterine contractions. *Med Eng Phys.* 2013;35:524–31.
- Childbirth simulator. <https://caehealthcare.com/patient-simulation/lucina>. Accessed 8 June 2017.
- Marieb EN, Hoehn K. Pregnancy and Human Development. In: Human anatomy and physiology. 10th ed: Pearson Education, Limited; 2015. p. 1116–7.
- Zhang J, Landy H, Branch D, Burkman R. Contemporary patterns of spontaneous labor with normal neonatal outcomes. *Obstet Gynecol.* 2013;116:1281–7.
- Fuchs A-R, Romero R, Keefe D, Parra M, Oyarzun E, Behnke E. Oxytocin secretion and human parturition: pulse frequency and duration increase during spontaneous labor in women. *Am J Obstet Gynecol.* 1991;165:1515–23.
- Becker KL. Physiology of vasopressin, oxytocin, and thirst. In: Principles and practice of endocrinology and metabolism. 3rd ed: Lippincott Williams and Wilkins; 2001. p. 276–84.
- Brandis K. Fluid Physiology. http://www.anaesthesiamcq.com/FluidBook/fl2_1.php. Accessed 1 June 2017.
- Caldeyro-Barcia R, Poseiro JJ. Oxytocin and contractility of the pregnant human uterus. *Ann N Y Acad Sci.* 1959;75(2 The Uterus):813–30.
- Antonucci MC, Pitman MC, Eid T, Steer PJ, Genevier ES. Simultaneous monitoring of head-to-cervix forces, intrauterine pressure and cervical dilatation during labour. *Med Eng Phys.* 1997;19:317–26.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

