

LETTER TO THE EDITOR

Open Access



A novel human peristalsis-inspired 3D-printed gastroduodenal simulator to evaluate intragastric/duodenal metabolic devices: a proof-of-concept study

Jinhee Kwon^{1,2}, Chang Seok Bang³, Sung Ock Kim⁴ and Do Hyun Park^{5*}

Keywords: Obesity, Gastric emptying, Endoscopic bariatric and metabolic therapies, 3D-printed gastroduodenal simulator

To the Editor,

Endoscopic bariatric and metabolic therapies (EBMT) is an attractive alternative to medical treatment and bariatric surgery owing to the efficacy and minimal invasiveness. Delay in gastric emptying or gastric retention with filling of the ingested food in stomach triggers satiation, and rapid gastric emptying is related to obesity [1, 2]. Therefore, delayed gastric emptying should be evaluated with an intragastric/duodenal device for EBMT to assess its performance in inducing weight loss in patients with obesity. However, no quantitative comparative results on the performance of delayed gastric emptying in intragastric balloons (IGBs), and gastro-duodenal flow restrictors (G-DFR) have been reported in a pre-clinical study. The aim of the present study was to introduce the 3D-printed gastroduodenal simulator that mimics human peristalsis as a virtual testing system for the quantitative measurement of delayed gastric emptying in novel intragastric/duodenal EBMT devices before animal trials can be conducted. We also planned a pilot animal study whether this novel device with substantial delayed gastric emptying in

experimental porcine group may affect relative weight loss compared with control porcine group.


The geometry and dimensions of the gastroduodenal simulator were designed to match the shape of the human stomach, based on computed tomography gastrography images of a healthy volunteer from a previous study [3]. The simulator was fabricated using 3D-printed molds (Omg SLA; Xiamen, China), elastomer (Ecoflex 00-30; Smooth-on, Inc., USA), and silicone adhesive (Sil-Poxy; Smooth-on, Inc.). The simulator comprised four segments (fundus, body, antrum, and pylorus-duodenal bulb), each of which was paired with two actuators (L16-R Miniature Linear Servos; Actuonix, British Columbia, Canada) to contact the stomach body. To simulate peristalsis, eight motors moving in a straight line were mounted to generate motility in these regions. The contraction wave (α), which controls the stroke parameter of peristalsis, was defined as follows: $\alpha = \text{diameter in contraction} / \text{diameter in relaxation}$. Each region was arranged with a relative contraction wave (α) of 0.9, 0.6, 0.6, and 0.2, respectively. The frequency parameter of peristalsis was set to twice per minute. Measurements in the simulator with peristalsis were obtained at different stages of drained and residual gastric material to assess gastric emptying volume defined as the partly drained fluid mass collected and retention ratio measured residual markers in the stomach. The normal range of values for gastric retention and gastric emptying after

*Correspondence: dhpark@amc.seoul.kr

⁵ Digestive Diseases Research Center, Department of Internal Medicine, University of Ulsan College of Medicine, Asan Medical Center, 88, Olympic-ro 43-gil, Songpa-gu, Seoul 05505, Republic of Korea
Full list of author information is available at the end of the article




© The Author(s) 2022. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.



(See figure on next page.)

Fig. 2 **A** Endoscopic images of the gastric duodenal simulator (upper-left) and a porcine stomach after deployment of the novel gastro-duodenal flow restrictor (G-DFR) (upper-middle). The pig model used two hemoclips (arrowheads) to anchor the proximal flap of the G-DFR (upper-right). Radiographs showing the G-DFR in a simulator with peristalsis (lower-left) and the porcine stomach (lower-middle) after deployment of the novel G-DFR. Injected contrast slowly ran off through a 60-mm distal PTFE skirt of G-DFR (lower-right, arrowheads). **B** Line graphs showing the differences in relative weight loss percentages between the control and experimental porcine groups with the placement of a G-DFR with a 60-mm distal PTFE skirt (left). Bar graphs representing the difference of body weight between the control and the experimental group at each time point (right)



2 h following a liquid meal were obtained from a prior study of 10 asymptomatic healthy volunteers in which it was defined as $25.9\% \pm (2 \times 12.5\%)$ [4]. This was consistent with the retention ratio of 28.61% we observed in our control experiment using the simulator with peristalsis for 4 h. (Additional file 1: Video S1).

The gastric emptying volume of experimental model occupying G-DFR with 60-mm polytetrafluoroethylene (PTFE) skirt was the least gastric emptying volume suggesting enhanced gastric retention (percentage of emptying volume: 32.4%, retention: 49.2%) compared with IGB (percentage of emptying volume: 47.9%, retention: 37.9%), and G-DFR with a 30-mm PTFE skirt (percentage of emptying volume: 41.7%, retention: 35.0%). (Fig. 1) To clarify the quantitative measurement of delayed gastric emptying with favorable results on G-DFR with 60 mm distal PTFE skirt in this simulator with peristalsis, exploratory porcine study was performed. (Fig. 2A) This pilot porcine study was only evaluated the short-term outcomes (28-day and 42-day follow up) in terms of body weight and G-DFR migration. In the exploratory porcine (Yorkshire pig, 35–40 kg, 2 in experimental and 2 in control group) study, experimental porcine group with a novel G-DFR showed relative weight loss relative to the control group but without statistical significance: -0.7% vs. 1.4% in 3 days; 5.2% vs. 15.7% in 14 days; 16.3% vs. 32.9% in 28 days; and 25.3% vs. 48.0% in 42 days. (Fig. 2B) No proximal or distal migration of a novel G-DFR was observed in the experimental group. (Additional file 2: Video S2).

In summary, we could quantitatively define the delayed gastric emptying volume for intragastric and gastroduodenal metabolic devices and evaluate the differences between these devices, which led to a delayed gastric emptying. The promising relative weight loss of our preliminary porcine study may suggest that this 3D-printed peristaltic simulator platform may predict the performance of a novel G-DFR with optimal delayed gastric emptying for EBMT.

Abbreviations

EBMT: Endoscopic bariatric and metabolic therapies; IGB: Intragastric balloon; G-DFR: Gastro-duodenal flow restrictors; PTFE: Polytetrafluoroethylene.

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s12967-022-03357-z>.

Additional file 1: Video S1. The methods and results of the experimental investigation in a simulator with peristalsis.

Additional file 2: Video S2. The methods and results of the exploratory porcine (Yorkshire pig, 35-40kg, 2 in experimental and 2 in control group) study.

Acknowledgements

Not applicable.

Authors' contributions

JK and DHP conceived the study design. DHP acquired financial support for the study leading to this publication. JK modeled and fabricated the gastro-duodenal simulator. SOK recorded the X-ray video and contributed to data acquisition. JK and SOK performed the experiments. JK, CSB, and DHP interpreted the data and performed the statistical analysis. CSB and DHP checked the integrity of the manuscript, including consistencies in the analysis results and data interpretation. JK and DHP wrote the first draft. CSB and DHP critically revised the draft. All authors have reviewed and approved the final draft for publication.

Funding

This work was supported by a Korea Medical Device Development Fund grant of the Korean government (the Ministry of Science and ICT, the Ministry of Trade, Industry and Energy, the Ministry of Health & Welfare, the Ministry of Food and Drug Safety) (Project Number: KMDF_PR_20200901_0266, 1711138598), by Basic Science Research Program through the National Research Foundation of Korea (NRF) funded by the Ministry of Education (2021R1A6A1A03040260) and by a grant (2021IP0015) from the Asan Institute for Life Sciences, Asan Medical Center, Seoul, Korea.

Availability of data and materials

The datasets used and/or analyzed during the current study are available from the corresponding author on reasonable request.

Declarations

Ethics approval and consent to participate

This study was approved by the Institutional Animal Care and Use Committee (IACUC no. 2021-12-049) as it accorded with its guidelines the humane handling of laboratory animals.

Consent for publication

Not applicable.

Competing interests

Do Hyun Park is a listed inventor on an issued patent for a gastric-duodenal flow restrictor (G-DFR) owned by the University of Ulsan Foundation for Industry Cooperation and the Asan Foundation. The authors have no financial relationships with any commercial entity producing healthcare-related products and/or services relevant to this article.

Author details

¹Division of Gastroenterology, Department of Internal Medicine, University of Ulsan College of Medicine, Asan Medical Center, Seoul, Republic of Korea. ²Department of Medical Science, Asan Medical Institute of Convergence Science and Technology, Asan Medical Center, University of Ulsan College of Medicine, Seoul, Republic of Korea. ³Department of Internal Medicine, Hallym University Chuncheon Sacred Heart Hospital, Chuncheon, Republic of Korea. ⁴Department of Radiology, University of Ulsan College of Medicine, Asan Medical Center, Seoul, Republic of Korea. ⁵Digestive Diseases Research Center, Department of Internal Medicine, University of Ulsan College of Medicine, Asan Medical Center, 88, Olympic-ro 43-gil, Songpa-gu, Seoul 05505, Republic of Korea.

Received: 22 March 2022 Accepted: 23 March 2022

Published online: 01 April 2022

References

1. Wright RA, Krinsky S, Fleeman C, Trujillo J, Teague E. Gastric emptying and obesity. *Gastroenterology*. 1983;84(4):747–51.
2. Abu Dayyeh BK, Acosta A, Camilleri M, et al. Endoscopic sleeve gastroplasty alters gastric physiology and induces loss of body weight in obese individuals. *Clin Gastroenterol Hepatol*. 2017;15(1):37–43.e31.
3. Kwon J, Choi J, Lee S, et al. Modelling and manufacturing of 3D-printed, patient-specific, and anthropomorphic gastric phantoms: a pilot study. *Sci Rep*. 2020;10(1):18976.

4. Maetani I, Ukita T, Tada T, et al. Gastric emptying in patients with palliative stenting for malignant gastric outlet obstruction. *Hepatogastroenterology*. 2008;55(81):298–302.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

